

primediagnosics

Protocol

Direct immuno fluorescence (IF) protocol

Our reagents are optimized for use in IF using object glasses with 4 mm diameter wells (Nutacon 10-342) and operating with a working volume of 5 μ l per well.

Due to differences in working conditions and circumstances, the working dilution of the specific antibody can vary between 5 and 1000 times. The appropriate working dilution should be determined on-site after receipt of the antibody.

The incubations are performed in a tightly closed humid box. For washing: gently cover the glasses completely with washing solution, leave for 5 min and remove the washing solution gently by tilting of the glass. Repeat this 3 – 5 times. Gently cover the glasses completely with demineralized water, leave for 1 min and remove the water by tilting the glass. Dry the glass on a heating plate at 50°C.

Buffers, chemicals & equipment:

- PBS 0.01 M: 8.18 gr NaCl, 0.15 gr KCl, 0.27 gr KH_2PO_4 , 1.42 gr $\text{Na}_2\text{HPO}_4 \times 2 \text{H}_2\text{O}$, pH 7.4
Add demineralized water to 1000 ml total volume
- Washing buffer: 0.001 M PBS (dilute PBS 10 times)
- Specific conjugate: FITC labelled polyclonal IgG. Working dilution is dependent on local circumstances and must be determined before usage with sample material. Working dilution may range from 5 to 500 times.
- Mounting-buffer: Vectashield mounting fluid (Brunschiwig H1000)
- Mounting-Kpi: 3.2 gr $\text{Na}_2\text{HPO}_4 \times 12 \text{H}_2\text{O}$, 0.15 gr $\text{NaH}_2\text{PO}_4 \times 2 \text{H}_2\text{O}$, pH 7.6
Add demineralized water to 100 ml total volume and add 50 ml glycerol
- Technical alcohol 96%
- Heating plate 50°C
- Incubation chamber 27°C
- Fluorescence microscope

Preparations:

The optimal working dilution of the specific conjugate is dependent on local circumstances. We strongly recommend to perform a titer test before starting an experiment with sample material:

1. Prepare a dilution series of a positive control, preferable a pure culture of the target bacteria: 10^7 , 10^6 and 10^5 cells/ml.
2. Prepare the IF-glasses as mentioned in step 1 of the protocol below.
3. Spot the bacterial cells according to the scheme below and follow step 2 of the protocol below.
4. Prepare a dilution series of the specific conjugate in PBS: 5, 50, 100, 250 and 500 times. If you have more wells you may extend the dilution series.
5. Spot the conjugate on the IF glasses according to the scheme below.
6. Follow steps 4 to 8 of the protocol below.
7. The optimal working dilution shows a fluorescence score 3-4 in all the bacterial concentrations¹.

	5	25	50	100	250	500	Dilution conjugate
10^5	0	0	0	0	0	0	
10^6	0	0	0	0	0	0	
10^7	0	0	0	0	0	0	

Bacterial cells/ml

Wageningen Plant Research

PrimeDiagnostics

PO Box 16, NL-6700 AA Wageningen, The Netherlands

Telephone: +31 (0)317 480 613

E-mail: primediagnosics@wur.nl

Website: www.primediagnosics.com

Protocol:

1. Clean the object glasses extensively with alcohol before use.
2. Spot the samples and controls on the glasses; allow air-drying on the heating plate and fixing the cells by flaming or 96% alcohol.
3. Dilute the specific conjugate to the optimal working dilution in PBS and spot the appropriate amount per well.
4. Incubate the glasses during 30 minutes at 27°C.
5. Wash the glasses and allow air-drying on the heating plate.
6. Spread the vectashield or mounting-Kpi in droplets over the glasses and cover this with a fitting standard object glass.
7. Examine the wells with 490 nm (blue light) and look for green fluorescent cells with the same morphology as the positive control. Use a magnification of 100 times.
8. Scoring of the results:
 - o 0 = no fluorescent cells visible
 - o 1 = weakly fluorescent cells visible
 - o 2 = moderately fluorescent cells visible
 - o 3 = clearly fluorescent cells visible
 - o 4 = bright fluorescent cells visible

Remarks:

1. When determining the optimal working dilution for the conjugate, please keep the following in mind that the fluorescence will be dimmed or disappear completely if:
 - The concentration of bacterial cells is too high.
 - The concentration of the conjugate is too high.
 - The concentration of the conjugate is too low.
 - The sample material contains auto-fluorescent particles.
 - The sample material contains an overflow of non-target bacterial cells.

Wageningen Plant Research

PrimeDiagnostics

PO Box 16, NL-6700 AA Wageningen, The Netherlands

Telephone: +31 (0)317 480 613

E-mail: primediagnostics@wur.nlWebsite: www.primediagnostics.com